



證書編號：L2773-160704

財團法人全國認證基金會  
Taiwan Accreditation Foundation

# 認 證 證 書

茲證明

億光電子工業股份有限公司

億光電子可靠度實驗室

新北市土城區中央路三段 76 巷 25 號

為本會認證之實驗室

認 證 依 據：ISO/IEC 17025：2005

認 證 編 號：2773

初 次 認 證 日 期：一百零二年七月五日

認 證 有 效 期 間：一百零五年七月五日至一百零八年七月四日止

認 證 範 圍：測試領域，如續頁

特 定 服 務 計 畫：美國能源之星實驗室認證服務計畫

董事長

陳 介 山

中華民國一百零五年七月四日



證書編號：L2773-160704

**財團法人全國認證基金會**  
**Taiwan Accreditation Foundation**

認證編號：2773

實驗室主管：戴億昌

18. 01 民生用品

LED 封裝, 模組, 陣列

0025 流明維持壽命測試

IES LM-80-08

光通量: 2.4 lm to 25000 lm

色坐標(CIE 1931 x/y): 0.00 to 1.00

電流範圍: 0 to 3 A(dc)

電壓範圍: 0 to 400 V(dc)

報告簽署人: 林煜騏; 戴億昌

**美國能源之星實驗室認證服務計畫**

18. 01 民生用品

LED 封裝, 模組, 陣列

0025 流明維持壽命測試

IES LM-80-08

光通量: 2.4 lm to 25000 lm

色坐標(CIE 1931 x/y): 0.00 to 1.00

電流範圍: 0 to 3 A(dc)

電壓範圍: 0 to 400 V(dc)

報告簽署人: 林煜騏; 戴億昌

(以下空白)



Certificate No. : L2773-160704

財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

This is to certify that

**Everlight Electronics Co., Ltd.**

**Reliability Lab, Everlight Electronics**

No.25, Ln. 76, Sec. 3, Zhongyang Rd., Tucheng Dist., New Taipei City 236, Taiwan (R.O.C.)

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2005  
**Accreditation Number** : 2773  
**Originally Accredited** : July 05, 2013  
**Effective Period** : July 05, 2016 to July 04, 2019  
**Accredited Scope** : Testing Field, see described in the Appendix  
**Specific Accreditation Program** : Accreditation Program for Laboratory of U.S. EPA  
ENERGY STAR



Jay-San Chen  
President, Taiwan Accreditation Foundation  
Date : July 04, 2016



Certificate No. : L2773-160704

**財團法人全國認證基金會**  
**Taiwan Accreditation Foundation**

Accreditation Number : 2773

Laboratory Head : TAI, Yi-Chang

18. 01 Commodity

LED package, module, array

O025 Lumen maintenance of life test

IES LM-80-08

Luminous flux: 2.4 lm to 25000 lm

Color coordinates (CIE 1931 x/y): 0.00 to 1.00

Current range: 0 to 3 A(dc)

Voltage range: 0 to 400 V(dc)

Approval Signatory: LIN, Yu-Chi; TAI, Yi-Chang

**Accreditation Program for Laboratory of U.S. EPA ENERGY STAR**

18. 01 Commodity

LED package, module, array

O025 Lumen maintenance of life test

IES LM-80-08

Luminous flux: 2.4 lm to 25000 lm

Color coordinates (CIE 1931 x/y): 0.00 to 1.00

Current range: 0 to 3 A(dc)

Voltage range: 0 to 400 V(dc)

Approval Signatory: LIN, Yu-Chi; TAI, Yi-Chang

(Null below)